
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002602
8:37 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:57
STERILIZING 3:15
136.0 C 220kPa 3:15
136.4 C 225kPa 3:45
136.4 C 225kPa 4:15
136.4 C 225kPa 4:45
136.3 C 224kPa 5:15
136.4 C 224kPa 5:45
136.4 C 225kPa 6:15
136.4 C 224kPa 6:45
136.4 C 224kPa 6:45
VENTING 6:45
AIR DRYING 7:39
CYCLE COMPLETE 30:02

Digital Signature #
C3FA3A422950ACCF
