
Midmark M11 - v1.0.5
Total Cycles: 3069

Name: RITTER2
Sterilizer ID: V2313228
11/01/2023 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.5 F 0.0
2:00 161.0 F 1.3
4:00 178.0 F 4.8
6:00 197.8 F 9.3
8:00 215.6 F 9.1
10:00 224.1 F 8.9
12:00 233.7 F 10.9
14:00 241.0 F 11.7
16:00 239.7 F 10.4
18:00 252.7 F 17.0
20:00 264.2 F 23.7
22:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.2 F 29.4
2:00 271.1 F 28.7
3:00 272.0 F 29.6
4:00 273.0 F 30.1

Min 271.0 F 28.7
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 235.7 F 0.0
10:00 200.5 F 0.0
15:00 192.9 F 0.0
20:00 189.0 F 0.0
25:00 186.6 F 0.0
30:00 195.9 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:05
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

11/01/2023 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____