
Midmark M11 - v1.0.0
Total Cycles: 3183

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2023 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 172.2 F 3.1
4:00 183.2 F 7.3
6:00 210.3 F 12.3
8:00 221.6 F 7.6
10:00 234.8 F 12.9
12:00 236.3 F 10.2
14:00 234.6 F 8.1
16:00 248.6 F 15.0
18:00 260.6 F 21.8
20:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.5
2:00 271.3 F 29.1
3:00 272.3 F 29.6
4:00 271.4 F 28.9

Min 271.2 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 214.0 F 0.0
10:00 232.8 F 0.0
15:00 235.5 F 0.0
20:00 228.4 F 0.0
25:00 224.5 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

05/01/2023 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____