
Midmark M11 - v1.0.5
Total Cycles: 662

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 181.5 F 2.6
4:00 188.0 F 7.3
6:00 211.0 F 12.9
8:00 224.0 F 8.0
10:00 238.6 F 13.5
12:00 240.2 F 11.8
14:00 240.7 F 11.1
16:00 254.4 F 18.1
18:00 265.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.1 F 29.3
2:00 272.0 F 29.8
3:00 271.7 F 29.1
4:00 272.4 F 29.8

Min 271.2 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 232.6 F 0.0
10:00 200.7 F 0.0
15:00 198.1 F 0.0
20:00 196.4 F 0.0
25:00 198.9 F 0.0
30:00 205.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

12/29/2021 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____