
Midmark M11 - v1.0.5
Total Cycles: 3965

Name: RITTER2
Sterilizer ID: V2313228
09/20/2024 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 169.6 F 3.0
4:00 192.7 F 7.7
6:00 210.5 F 13.2
8:00 224.2 F 7.5
10:00 237.8 F 12.9
12:00 239.5 F 10.8
14:00 237.9 F 9.7
16:00 252.1 F 16.7
18:00 264.2 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.2 F 29.0
2:00 272.6 F 29.5
3:00 271.8 F 28.8
4:00 272.5 F 29.1

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 226.1 F 0.0
10:00 208.1 F 0.0
15:00 207.2 F 0.0
20:00 207.4 F 0.0
25:00 205.7 F 0.0
30:00 203.6 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

09/20/2024 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____