
Midmark M11 - v1.0.0
Total Cycles: 795

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2021 02:06 PM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 177.8 F 3.2
4:00 203.3 F 4.0
6:00 218.3 F 5.0
8:00 224.0 F 4.0
10:00 229.1 F 5.5
12:00 237.9 F 9.4
14:00 250.8 F 15.6

STERILIZING
mm:ss Degrees PSI
0:00 251.6 F 15.9
1:00 252.2 F 16.1
2:00 252.9 F 16.5
3:00 251.7 F 15.9
4:00 253.0 F 16.5
5:00 251.9 F 15.9
6:00 252.9 F 16.6
7:00 252.3 F 16.2
8:00 251.7 F 15.9
9:00 252.9 F 16.6
10:00 252.6 F 16.3
11:00 252.0 F 16.0
12:00 251.8 F 16.0
13:00 253.0 F 16.6
14:00 252.7 F 16.4
15:00 252.3 F 16.2
16:00 251.8 F 15.9
17:00 252.0 F 16.2
18:00 252.9 F 16.6
19:00 252.8 F 16.4
20:00 252.4 F 16.2
21:00 252.0 F 16.0
22:00 251.7 F 15.9
23:00 252.5 F 16.5
24:00 252.9 F 16.6
25:00 252.8 F 16.4
26:00 252.4 F 16.3
27:00 252.1 F 16.1
28:00 251.8 F 15.9
29:00 251.9 F 16.2
30:00 252.9 F 16.6

Min 251.6 F 15.8
Max 253.1 F 16.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 205.4 F 0.0
10:00 193.3 F 0.0
15:00 166.1 F 0.0
20:00 158.5 F 0.0
Dry Cycle Aborted

FILLING: 2:10
HEATING: 14:40
STERILIZING: 30:00
VENTING: 2:57
DRYING: 22:23

TOTAL CYCLE: 01:12:09

04/21/2021 03:18 PM

ACCEPTED: X

REJECTED:

Operator: U01