

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2480

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2022 03:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 176.6 F 3.5  
4:00 193.5 F 8.0  
6:00 216.9 F 13.3  
8:00 223.7 F 6.9  
10:00 235.9 F 11.9  
12:00 235.2 F 8.7  
14:00 240.3 F 10.2  
16:00 246.6 F 13.4  
18:00 258.5 F 20.1  
20:00 269.5 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.6 F 28.4  
3:00 272.3 F 29.3  
4:00 272.1 F 28.8  
  
Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 209.8 F 0.0  
10:00 229.0 F 0.0  
15:00 235.1 F 0.0  
20:00 229.5 F 0.0  
25:00 226.9 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:05  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

10/24/2022 04:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_