
Midmark M11 - v1.0.5
Total Cycles: 4424

Name: RITTER2
Sterilizer ID: V2313228
03/13/2025 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.3 F 0.0
2:00 144.6 F 1.3
4:00 147.4 F 3.5
6:00 176.5 F 7.2
8:00 195.4 F 10.8
10:00 222.4 F 5.5
12:00 221.7 F 8.8
14:00 234.5 F 13.3
16:00 236.0 F 10.1
18:00 238.0 F 9.1
20:00 246.1 F 13.9
22:00 256.9 F 19.8
24:00 266.5 F 26.0
26:00 271.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.2
1:00 271.6 F 30.5
2:00 271.6 F 30.5
3:00 271.9 F 30.3
4:00 271.8 F 30.2

Min 271.5 F 30.1
Max 272.5 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 206.7 F 0.0
10:00 205.5 F 0.0
15:00 200.3 F 0.0
20:00 196.8 F 0.0
25:00 196.9 F 0.0
30:00 196.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:54
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:05:49
CYCLE COMPLETE

03/13/2025 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____