
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002165
10:06 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:42
136.4 C 226kPa 3:42
136.1 C 221kPa 4:12
136.3 C 224kPa 4:42
136.3 C 225kPa 5:12
136.4 C 226kPa 5:42
136.4 C 225kPa 6:12
136.1 C 222kPa 6:42
136.2 C 223kPa 7:12
136.2 C 223kPa 7:12
VENTING 7:12
AIR DRYING 8:04
CYCLE COMPLETE 18:18

Digital Signature #
6B94CD5708183B
