
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002896
10:31 29/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:30
PRESSURIZING 1:38
STERILIZING 2:28
136.0 C 220kPa 2:28
136.2 C 224kPa 2:58
136.4 C 223kPa 3:28
136.4 C 224kPa 3:58
136.4 C 220kPa 4:28
136.4 C 222kPa 4:58
136.4 C 225kPa 5:28
136.4 C 224kPa 5:58
136.4 C 224kPa 5:58
VENTING 5:58
AIR DRYING 6:50
CYCLE COMPLETE 6:53

Digital Signature #
7C325D7A8F04F9CE
