

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003619  
8:38 09/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:19  
STERILIZING 3:21  
136.4 C 222kPa 3:21  
136.3 C 223kPa 3:51  
136.4 C 223kPa 4:21  
136.1 C 217kPa 4:51  
136.1 C 217kPa 5:21  
136.2 C 222kPa 5:51  
136.4 C 223kPa 6:21  
136.1 C 221kPa 6:51  
136.1 C 221kPa 6:51  
VENTING 6:51  
AIR DRYING 7:41  
CYCLE COMPLETE 7:55

-----  
Digital Signature #  
FD8C531693C625F8  
-----