
Midmark M11 - v1.0.5
Total Cycles: 534

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 186.1 F 3.2
4:00 186.3 F 7.4
6:00 203.7 F 12.3
8:00 218.0 F 6.4
10:00 231.2 F 11.4
12:00 232.9 F 8.6
14:00 240.9 F 11.0
16:00 245.4 F 13.4
18:00 257.1 F 19.7
20:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 271.9 F 29.5
2:00 272.1 F 29.2
3:00 271.4 F 29.1
4:00 272.5 F 29.7

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 218.3 F 0.0
10:00 228.4 F 0.0
15:00 226.1 F 0.0
20:00 225.6 F 0.0
25:00 225.9 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

11/10/2021 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____