

-----  
Midmark M11 - v1.0.5  
Total Cycles: 367

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2021 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 179.2 F 3.7  
4:00 185.6 F 9.5  
6:00 212.0 F 6.5  
8:00 225.1 F 12.2  
10:00 233.2 F 10.8  
12:00 237.9 F 10.1  
14:00 251.2 F 17.5  
16:00 265.6 F 25.1  
18:00 271.6 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 271.9 F 30.0  
2:00 272.4 F 30.4  
3:00 272.6 F 30.6  
4:00 272.6 F 30.7  
  
Min 271.6 F 29.5  
Max 274.0 F 31.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.7 F 0.0  
5:00 221.9 F 0.0  
10:00 229.8 F 0.0  
15:00 228.0 F 0.0  
20:00 223.3 F 0.0  
25:00 222.4 F 0.0  
30:00 190.6 F 0.0  
Drying Complete

FILLING: 2:08  
HEATING: 18:02  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:57:09  
CYCLE COMPLETE

09/15/2021 11:34 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01