
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006769
8:32 20/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:52
136.5 C 222kPa 4:52
135.8 C 220kPa 5:22
136.4 C 219kPa 5:52
136.2 C 219kPa 6:22
136.4 C 222kPa 6:52
136.3 C 222kPa 7:22
136.1 C 220kPa 7:52
136.4 C 223kPa 8:22
136.4 C 223kPa 8:22
VENTING 8:22
AIR DRYING 9:13
CYCLE COMPLETE 10:50

Digital Signature #
77F252B987D7EEB4
