
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005276
14:26 27/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:55
STERILIZING 4:00
136.4 C 220kPa 4:00
136.4 C 225kPa 4:30
136.0 C 221kPa 5:00
136.3 C 223kPa 5:30
136.3 C 222kPa 6:00
136.2 C 219kPa 6:30
136.3 C 223kPa 7:00
136.2 C 223kPa 7:30
136.2 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:23
CYCLE COMPLETE 15:49

Digital Signature #
ECFCB95CEA9BCCD5
