
Midmark M11 - v1.0.0
Total Cycles: 4504

Name: RITTER 1
Sterilizer ID: V2178198
04/30/2024 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 73.4 F 0.1
4:00 84.9 F 0.6
6:00 103.7 F 2.4
8:00 136.0 F 4.7
10:00 162.6 F 7.2
12:00 182.0 F 10.2
14:00 200.3 F 11.8
16:00 208.3 F 4.5
18:00 218.4 F 7.8
20:00 229.0 F 11.3
22:00 227.8 F 6.3
24:00 239.4 F 11.3
26:00 235.2 F 8.5
28:00 248.1 F 14.3
30:00 258.4 F 19.9
32:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.6 F 29.1
4:00 271.8 F 28.3

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 201.9 F 0.0
10:00 193.9 F 0.0
15:00 215.1 F 0.0
20:00 221.5 F 0.0
25:00 221.8 F 0.0
30:00 224.9 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 33:26
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:11:41
CYCLE COMPLETE

04/30/2024 10:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____