

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2393

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2022 08:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
2:00 205.1 F 2.6  
4:00 203.9 F 7.3  
6:00 223.4 F 13.4  
8:00 231.6 F 8.9  
10:00 238.5 F 9.2  
12:00 243.7 F 12.8  
14:00 242.6 F 11.5  
16:00 256.6 F 18.8  
18:00 268.7 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.5 F 28.9  
2:00 272.1 F 28.9  
3:00 271.7 F 28.3  
4:00 272.6 F 29.0  
  
Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 233.5 F 0.0  
10:00 228.0 F 0.0  
15:00 227.7 F 0.0  
20:00 226.1 F 0.0  
25:00 227.1 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:57:38  
CYCLE COMPLETE

10/04/2022 09:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_