

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 006186  
15:10 28/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:57  
STERILIZING 3:55  
136.0 C 220kPa 3:55  
135.9 C 221kPa 4:25  
136.4 C 224kPa 4:55  
136.1 C 220kPa 5:25  
136.3 C 224kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 224kPa 6:55  
136.2 C 219kPa 7:25  
136.2 C 219kPa 7:25  
VENTING 7:25  
AIR DRYING 8:21  
CYCLE COMPLETE 12:17

-----

Digital Signature #  
DED270B4CE865E5D  
-----