
Midmark M11 - v1.0.5
Total Cycles: 1711

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 04:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 193.9 F 0.7
4:00 206.0 F 5.5
6:00 229.1 F 11.4
8:00 230.6 F 7.9
10:00 241.7 F 11.9
12:00 243.8 F 12.9
14:00 244.3 F 12.6
16:00 257.9 F 20.0
18:00 269.8 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.6
2:00 271.8 F 28.8
3:00 271.9 F 29.4
4:00 272.4 F 29.4

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 224.5 F 0.0
10:00 239.7 F 0.0
15:00 233.0 F 0.0
20:00 220.5 F 0.0
25:00 219.5 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

11/10/2022 05:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____