
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001209
11:05 11/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:35
STERILIZING 4:42
136.0 C 220kPa 4:42
136.1 C 223kPa 5:12
136.0 C 221kPa 5:42
136.4 C 223kPa 6:12
136.2 C 222kPa 6:42
136.3 C 223kPa 7:12
136.4 C 222kPa 7:42
136.4 C 223kPa 8:12
136.4 C 223kPa 8:12
VENTING 8:12
AIR DRYING 9:08
CYCLE COMPLETE 11:22

Digital Signature #
B4D0B2BCC1328D27
