
Midmark M11 - v1.0.0
Total Cycles: 4253

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2024 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.1 F 0.0
2:00 103.9 F 0.9
4:00 122.2 F 3.0
6:00 159.7 F 6.0
8:00 183.4 F 9.9
10:00 209.6 F 6.1
12:00 217.8 F 5.7
14:00 224.6 F 9.5
16:00 235.2 F 12.1
18:00 237.4 F 9.6
20:00 240.1 F 10.1
22:00 245.6 F 12.8
24:00 257.2 F 19.0
26:00 267.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.8 F 28.2
2:00 272.1 F 28.4
3:00 272.5 F 28.6
4:00 272.4 F 28.9

Min 271.3 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 235.6 F 0.0
10:00 229.5 F 0.0
15:00 227.3 F 0.0
20:00 221.8 F 0.0
25:00 222.2 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:26
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:05:59
CYCLE COMPLETE

01/31/2024 02:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____