
Midmark M11 - v1.0.0
Total Cycles: 2916

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 150.5 F 3.1
4:00 170.4 F 6.7
6:00 193.7 F 11.1
8:00 211.9 F 4.3
10:00 223.9 F 9.4
12:00 232.6 F 9.4
14:00 238.4 F 11.0
16:00 235.6 F 8.8
18:00 249.9 F 15.4
20:00 261.5 F 21.9
22:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.3
2:00 272.3 F 28.9
3:00 271.4 F 28.4
4:00 271.9 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 212.7 F 0.0
10:00 214.2 F 0.0
15:00 213.1 F 0.0
20:00 218.6 F 0.0
25:00 225.3 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

02/16/2023 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____