

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002106  
10:11 02/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:05  
STERILIZING 3:13  
136.2 C 223kPa 3:13  
136.4 C 225kPa 3:43  
136.4 C 222kPa 4:13  
136.4 C 225kPa 4:43  
136.4 C 225kPa 5:13  
136.4 C 225kPa 5:43  
136.3 C 225kPa 6:13  
136.1 C 222kPa 6:43  
136.1 C 222kPa 6:43  
VENTING 6:43  
AIR DRYING 7:34  
CYCLE COMPLETE 37:34  
-----

Digital Signature #  
C8FBDDC261E4AC23  
-----