
Midmark M11 - v1.0.5
Total Cycles: 3949

Name: RITTER2
Sterilizer ID: V2313228
09/16/2024 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.0 F 0.0
2:00 87.5 F 0.0
4:00 94.7 F 0.0
6:00 130.9 F 0.1
8:00 169.8 F 0.2
10:00 186.8 F 3.2
12:00 199.9 F 6.4
14:00 211.2 F 10.0
16:00 226.1 F 10.4
18:00 230.2 F 8.9
20:00 240.8 F 13.5
22:00 240.1 F 11.3
24:00 237.9 F 9.8
26:00 250.8 F 16.0
28:00 261.8 F 22.3
30:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.5
2:00 271.8 F 28.7
3:00 272.0 F 29.4
4:00 272.4 F 29.3

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 203.2 F 0.0
10:00 202.9 F 0.0
15:00 202.1 F 0.0
20:00 206.2 F 0.0
25:00 209.6 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

09/16/2024 04:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____