
Midmark M11 - v1.0.5
Total Cycles: 582

Name: RITTER2
Sterilizer ID: V2313228
11/29/2021 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 174.6 F 4.0
4:00 196.3 F 8.9
6:00 213.7 F 7.8
8:00 230.4 F 9.9
10:00 232.4 F 8.4
12:00 243.0 F 12.0
14:00 245.6 F 13.4
16:00 258.6 F 20.5
18:00 269.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.3 F 29.7
2:00 271.4 F 28.8
3:00 272.3 F 29.6
4:00 271.6 F 28.8

Min 271.3 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 236.3 F 0.0
10:00 199.3 F 0.0
15:00 190.9 F 0.0
20:00 184.4 F 0.0
25:00 190.9 F 0.0
30:00 180.1 F 0.0

Drying Complete

FILLING: 2:19
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

11/29/2021 01:53 PM
ACCEPTED: X
REJECTED:
Operator: U02