
Midmark M11 - v1.0.5
Total Cycles: 3550

Name: RITTER2
Sterilizer ID: V2313228
04/23/2024 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.4 F 0.0
2:00 85.6 F 0.4
4:00 102.1 F 1.8
6:00 127.5 F 4.1
8:00 148.7 F 6.8
10:00 171.6 F 10.1
12:00 213.3 F 10.1
14:00 209.7 F 6.6
16:00 225.3 F 10.3
18:00 230.2 F 8.6
20:00 236.6 F 11.0
22:00 233.9 F 8.6
24:00 247.2 F 14.6
26:00 257.4 F 20.3
28:00 267.2 F 26.5
30:00 273.1 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.1
1:00 271.8 F 30.1
2:00 271.8 F 30.0
3:00 272.0 F 30.0
4:00 272.4 F 30.2

Min 271.5 F 30.0
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 227.3 F 0.0
10:00 224.3 F 0.0
15:00 196.7 F 0.0
20:00 186.1 F 0.0
25:00 189.0 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 30:13
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

04/23/2024 03:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____