
Midmark M11 - v1.0.0
Total Cycles: 5279

Name: RITTER 1
Sterilizer ID: V2178198
01/27/2025 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.2 F 0.0
2:00 143.1 F 1.4
4:00 159.0 F 4.3
6:00 180.8 F 7.9
8:00 196.1 F 11.8
10:00 201.2 F 3.5
12:00 215.8 F 7.4
14:00 227.4 F 11.5
16:00 227.1 F 7.4
18:00 239.1 F 12.7
20:00 238.5 F 10.7
22:00 250.7 F 16.5
24:00 261.0 F 22.5
26:00 271.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.4 F 29.9
2:00 273.0 F 29.9
3:00 272.2 F 29.5
4:00 271.6 F 29.2

Min 271.4 F 29.1
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 205.2 F 0.0
10:00 189.4 F 0.0
15:00 186.7 F 0.0
20:00 193.9 F 0.0
25:00 199.9 F 0.0
30:00 207.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:50
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:05:36
CYCLE COMPLETE

01/27/2025 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____