
Midmark M11 - v1.0.0
Total Cycles: 5378

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2025 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 158.7 F 2.3
4:00 173.8 F 5.8
6:00 212.2 F 10.5
8:00 224.0 F 4.5
10:00 222.5 F 8.0
12:00 234.7 F 12.5
14:00 233.6 F 8.7
16:00 242.4 F 11.1
18:00 244.5 F 12.6
20:00 256.2 F 18.9
22:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.5 F 28.5
3:00 272.1 F 29.2
4:00 272.2 F 28.9

Min 271.1 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 198.4 F 0.0
10:00 186.3 F 0.0
15:00 186.3 F 0.0
20:00 184.5 F 0.0
25:00 183.4 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:30
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

02/28/2025 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____