
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000545
14:51 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:21
STERILIZING 4:28
136.0 C 220kPa 4:28
136.4 C 233kPa 4:58
136.3 C 232kPa 5:28
136.3 C 230kPa 5:58
136.2 C 231kPa 6:28
136.4 C 232kPa 6:58
136.2 C 231kPa 7:28
136.3 C 231kPa 7:58
136.3 C 231kPa 7:58
VENTING 7:58
AIR DRYING 8:54
CYCLE COMPLETE 22:01

Digital Signature #
2001D71DBE66035C
