
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001103
16:03 01/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:24
PRESSURIZING 4:25
STERILIZING 5:42
136.4 C 222kPa 5:42
136.4 C 226kPa 6:12
136.3 C 223kPa 6:42
136.4 C 225kPa 7:12
136.2 C 223kPa 7:42
136.4 C 222kPa 8:12
136.4 C 225kPa 8:42
136.4 C 225kPa 9:12
136.4 C 225kPa 9:12
VENTING 9:12
AIR DRYING 10:10
CYCLE COMPLETE 25:10

Digital Signature #
2473415094254885
