
Midmark M11 - v1.0.5
Total Cycles: 2235

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 04:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 178.9 F 3.7
4:00 196.0 F 8.1
6:00 215.1 F 13.0
8:00 227.4 F 8.9
10:00 235.2 F 10.3
12:00 241.7 F 12.6
14:00 241.8 F 11.8
16:00 255.0 F 18.7
18:00 266.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 29.7
2:00 271.2 F 29.1
3:00 272.5 F 29.8
4:00 271.6 F 29.1

Min 271.2 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 232.0 F 0.0
10:00 231.1 F 0.0
15:00 226.6 F 0.0
20:00 222.2 F 0.0
25:00 221.0 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

04/11/2023 05:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____