
Midmark M11 - v1.0.0
Total Cycles: 1129

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2021 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.8 F 0.0
2:00 146.5 F 1.2
4:00 176.8 F 4.8
6:00 194.8 F 9.1
8:00 213.2 F 7.3
10:00 228.3 F 7.6
12:00 239.3 F 12.5
14:00 236.1 F 9.3
16:00 238.1 F 9.3
18:00 247.0 F 13.8
20:00 259.6 F 20.7
22:00 270.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 271.9 F 28.9
2:00 272.7 F 29.1
3:00 271.6 F 28.5
4:00 271.3 F 28.6

Min 271.1 F 28.2
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 224.1 F 0.0
10:00 190.6 F 0.0
15:00 185.4 F 0.0
20:00 185.1 F 0.0
25:00 179.1 F 0.0
30:00 178.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

08/30/2021 11:19 AM
ACCEPTED: X
REJECTED:
Operator: U02