
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000972
9:45 14/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:19
STERILIZING 4:27
136.0 C 220kPa 4:27
136.3 C 221kPa 4:57
136.2 C 223kPa 5:27
136.4 C 224kPa 5:57
136.2 C 220kPa 6:27
136.4 C 221kPa 6:57
136.2 C 223kPa 7:27
136.4 C 223kPa 7:57
136.4 C 223kPa 7:57
VENTING 7:57
AIR DRYING 8:53
CYCLE COMPLETE 8:54

Digital Signature #
CDA85DCCBF5F2B82
