

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4098

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2024 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.1 F 0.0  
2:00 131.2 F 0.2  
4:00 159.0 F 2.9  
6:00 182.1 F 6.3  
8:00 200.5 F 10.0  
10:00 210.9 F 10.2  
12:00 225.1 F 7.7  
14:00 232.9 F 11.9  
16:00 233.0 F 8.8  
18:00 240.7 F 11.6  
20:00 244.3 F 12.9  
22:00 255.4 F 18.7  
24:00 265.6 F 24.9  
26:00 271.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.5  
1:00 272.4 F 29.5  
2:00 272.7 F 29.8  
3:00 272.2 F 30.0  
4:00 271.9 F 29.4

Min 271.8 F 29.4  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 228.9 F 0.0  
10:00 222.6 F 0.0  
15:00 225.7 F 0.0  
20:00 232.0 F 0.0  
25:00 230.8 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 26:39  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:07  
CYCLE COMPLETE

11/07/2024 04:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_