

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000765  
9:53 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:13  
STERILIZING 4:26  
136.5 C 233kPa 4:26  
136.3 C 233kPa 4:56  
136.4 C 233kPa 5:26  
136.3 C 233kPa 5:56  
136.3 C 231kPa 6:26  
136.4 C 233kPa 6:56  
136.4 C 233kPa 7:26  
136.4 C 232kPa 7:56  
136.4 C 232kPa 7:56  
VENTING 7:56  
AIR DRYING 8:51  
CYCLE COMPLETE 8:52

-----  
Digital Signature #  
DCEC7C3C5337F0D0  
-----