
Midmark M11 - v1.0.0
Total Cycles: 2553

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2022 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.8 F 0.0
2:00 203.1 F 4.4
4:00 216.5 F 9.3
6:00 226.0 F 7.1
8:00 235.5 F 9.8
10:00 235.0 F 8.0
12:00 245.3 F 12.8
14:00 242.6 F 11.3
16:00 255.1 F 18.1
18:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.6 F 29.1
2:00 271.6 F 28.4
3:00 272.5 F 29.2
4:00 272.0 F 28.6

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 232.1 F 0.0
10:00 240.0 F 0.0
15:00 234.6 F 0.0
20:00 226.8 F 0.0
25:00 224.5 F 0.0
30:00 225.6 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

11/14/2022 03:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____