

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1597

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2022 07:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.7 F 0.0  
2:00 76.9 F 0.0  
4:00 100.3 F 0.1  
6:00 125.2 F 2.3  
8:00 151.7 F 4.9  
10:00 173.7 F 7.8  
12:00 186.8 F 11.6  
14:00 206.3 F 2.7  
16:00 216.8 F 6.3  
18:00 227.2 F 10.0  
20:00 234.2 F 9.9  
22:00 237.9 F 10.3  
24:00 236.6 F 8.5  
26:00 247.0 F 13.6  
28:00 258.1 F 19.7  
30:00 268.3 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 29.0  
2:00 271.5 F 28.1  
3:00 272.6 F 29.1  
4:00 272.1 F 28.5

Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 216.3 F 0.0  
10:00 222.5 F 0.0  
15:00 225.1 F 0.0  
20:00 224.8 F 0.0  
25:00 227.8 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:09:26  
CYCLE COMPLETE

02/07/2022 09:00 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_