

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2970

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2023 03:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.1 F 0.0  
2:00 169.5 F 3.0  
4:00 194.6 F 7.3  
6:00 221.2 F 12.4  
8:00 227.1 F 7.8  
10:00 239.5 F 13.2  
12:00 239.3 F 10.9  
14:00 237.8 F 9.8  
16:00 251.8 F 16.5  
18:00 263.5 F 23.4  
20:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.2 F 29.3  
2:00 271.1 F 28.8  
3:00 272.8 F 30.1  
4:00 272.3 F 29.4  
  
Min 270.9 F 28.6  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
5:00 229.3 F 0.0  
10:00 220.8 F 0.0  
15:00 220.7 F 0.0  
20:00 186.9 F 0.0  
25:00 183.4 F 0.0  
30:00 180.4 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

10/04/2023 04:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_