
Midmark M11 - v1.0.0
Total Cycles: 2352

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2022 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 191.4 F 4.5
4:00 200.2 F 9.5
6:00 222.7 F 7.0
8:00 232.7 F 10.5
10:00 232.9 F 7.2
12:00 243.1 F 12.3
14:00 244.7 F 12.7
16:00 257.4 F 19.5
18:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.2 F 28.9
2:00 271.8 F 29.1
3:00 271.9 F 28.8
4:00 271.8 F 29.1

Min 271.3 F 28.5
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 228.4 F 0.0
10:00 234.6 F 0.0
15:00 237.7 F 0.0
20:00 222.8 F 0.0
25:00 229.7 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

09/21/2022 03:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____