
Midmark M11 - v1.0.0
Total Cycles: 3236

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 04:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 187.2 F 1.7
4:00 207.3 F 7.1
6:00 222.9 F 12.8
8:00 230.7 F 8.8
10:00 238.4 F 10.3
12:00 244.3 F 12.9
14:00 243.6 F 12.1
16:00 257.2 F 19.1
18:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 272.1 F 28.9
4:00 272.5 F 28.9

Min 271.3 F 28.1
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 222.2 F 0.0
10:00 237.5 F 0.0
15:00 229.8 F 0.0
20:00 228.1 F 0.0
25:00 226.2 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/15/2023 05:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____