

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000823  
9:43 17/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 3:05  
STERILIZING 4:18  
136.0 C 227kPa 4:18  
135.9 C 228kPa 4:48  
136.3 C 229kPa 5:18  
136.3 C 230kPa 5:48  
136.2 C 232kPa 6:18  
136.2 C 230kPa 6:48  
136.4 C 231kPa 7:18  
136.3 C 231kPa 7:48  
136.3 C 231kPa 7:48  
VENTING 7:49  
AIR DRYING 8:41  
CYCLE COMPLETE 23:41  
-----

Digital Signature #  
4A575FE1255819BB  
-----