
Midmark M11 - v1.0.0
Total Cycles: 1469

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2021 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 167.6 F 2.2
4:00 195.5 F 5.8
6:00 210.3 F 10.6
8:00 219.3 F 5.1
10:00 232.2 F 9.2
12:00 239.9 F 11.0
14:00 240.9 F 11.4
16:00 238.4 F 9.8
18:00 252.3 F 16.5
20:00 263.8 F 23.3
22:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.5 F 29.5
2:00 272.0 F 28.8
3:00 271.9 F 29.2
4:00 272.7 F 29.2

Min 271.2 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 233.0 F 0.0
10:00 222.4 F 0.0
15:00 209.8 F 0.0
20:00 207.2 F 0.0
25:00 198.6 F 0.0
Dry Cycle Aborted

FILLING: 1:30
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:22
DRYING: 28:12
TOTAL CYCLE: 00:59:47

12/16/2021 03:16 PM
ACCEPTED: X
REJECTED:
Operator: U02