
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001596
15:13 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 4:36
STERILIZING 6:06
136.0 C 220kPa 6:06
136.4 C 223kPa 6:36
136.2 C 221kPa 7:06
136.3 C 220kPa 7:36
136.4 C 219kPa 8:07
136.4 C 222kPa 8:37
136.4 C 222kPa 9:07
136.0 C 218kPa 9:36
VENTING 9:36
AIR DRYING 10:41
CYCLE COMPLETE 15:27

Digital Signature #
6913745FCEEBCD13
