

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000921  
12:33 10/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 3:48  
CYCLE ABORTED 9:48  
CYCLE FAULT 4  
NOT STERILE  
END TIME 9:48  
-----

Digital Signature #  
37AED391C343C7C9  
-----