

-----  
Midmark M11 - v1.0.0  
Total Cycles: 638

Name: MIDMARK  
Sterilizer ID: V2178198  
01/28/2021 07:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.6 F 0.0  
2:00 75.0 F 0.0  
4:00 79.1 F 0.0  
6:00 104.4 F 0.1  
8:00 139.8 F 0.2  
10:00 180.5 F 2.0  
12:00 200.0 F 5.3  
14:00 207.8 F 8.8  
16:00 220.6 F 12.5  
18:00 225.1 F 7.0  
20:00 235.7 F 11.8  
22:00 235.1 F 8.5  
24:00 240.8 F 10.4  
26:00 246.7 F 13.4  
28:00 259.2 F 20.1  
30:00 270.1 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.1 F 28.3  
2:00 272.2 F 28.8  
3:00 271.5 F 27.9  
4:00 272.4 F 28.7

Min 271.3 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 227.7 F 0.0  
10:00 230.0 F 0.0  
15:00 229.6 F 0.0  
20:00 225.8 F 0.0  
25:00 230.0 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:51  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:08:44  
CYCLE COMPLETE

01/28/2021 08:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_