
Midmark M11 - v1.0.5
Total Cycles: 1096

Name: RITTER2
Sterilizer ID: V2313228
05/11/2022 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 187.3 F 3.9
4:00 194.1 F 8.9
6:00 217.1 F 7.8
8:00 228.9 F 9.6
10:00 234.8 F 9.0
12:00 244.2 F 13.2
14:00 244.5 F 12.9
16:00 257.4 F 19.8
18:00 268.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.8 F 29.5
2:00 271.7 F 28.7
3:00 272.7 F 29.7
4:00 272.2 F 29.1

Min 271.5 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 210.1 F 0.0
10:00 223.8 F 0.0
15:00 226.3 F 0.0
20:00 226.6 F 0.0
25:00 218.6 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

05/11/2022 02:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____