
Midmark M11 - v1.0.5
Total Cycles: 2982

Name: RITTER2
Sterilizer ID: V2313228
10/05/2023 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C102: FILL MODE
STOP PRESSED
CHAMBER TEMP 172F

** ITEMS NOT STERILE **

10/05/2023 03:48 PM