
Midmark M11 - v1.0.5
Total Cycles: 2774

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 05:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 195.7 F 2.1
4:00 215.3 F 7.9
6:00 229.1 F 12.7
8:00 235.0 F 10.6
10:00 234.9 F 9.0
12:00 238.6 F 9.7
14:00 250.0 F 15.7
16:00 262.5 F 22.9
18:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.0 F 28.9
2:00 272.5 F 29.4
3:00 272.2 F 29.6
4:00 271.8 F 28.8

Min 271.6 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 230.0 F 0.0
10:00 233.7 F 0.0
15:00 227.4 F 0.0
20:00 223.7 F 0.0
25:00 224.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

08/15/2023 06:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____