
Midmark M11 - v1.0.5
Total Cycles: 961

Name: RITTER2
Sterilizer ID: V2313228
04/06/2022 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 179.9 F 0.2
4:00 185.4 F 1.6
6:00 208.8 F 6.5
8:00 222.1 F 11.9
10:00 230.0 F 8.3
12:00 240.6 F 11.7
14:00 243.6 F 12.8
16:00 243.6 F 12.4
18:00 256.8 F 19.4
20:00 268.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.7
2:00 272.3 F 29.1
3:00 271.4 F 28.9
4:00 272.6 F 29.8

Min 271.4 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 215.3 F 0.0
10:00 226.3 F 0.0
15:00 225.8 F 0.0
20:00 223.6 F 0.0
25:00 223.1 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

04/06/2022 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____