
Midmark M11 - v1.0.0
Total Cycles: 2543

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 169.9 F 4.1
4:00 182.3 F 8.4
6:00 202.9 F 13.4
8:00 219.6 F 7.2
10:00 233.7 F 12.4
12:00 235.0 F 9.3
14:00 237.7 F 9.4
16:00 246.9 F 13.9
18:00 259.0 F 20.6
20:00 270.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.7 F 29.3
2:00 271.7 F 28.5
3:00 272.6 F 29.2
4:00 271.6 F 28.5

Min 271.5 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 220.7 F 0.0
10:00 236.1 F 0.0
15:00 231.2 F 0.0
20:00 227.6 F 0.0
25:00 227.4 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

11/10/2022 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____